

Biolab SRL

MONTHLY IMMUNOASSAY END OF CYCLE REPORT

CYCLE 17

Authorised by: Stephen Doherty, RIQAS Manager

Issue No: 1

Issue Date: 04/02/2020

Randox Laboratories Limited
55 Diamond Road
CRUMLIN BT29 4QY
Tel: +44 (0)28 9445 4399
Fax: +44 (0)28 9445 4398
Email: mail@riqas.com





AFP, ng/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse GSeries
Reagent: Fujirebio Inc.

RIQAS TDPA: 13.7% **Biological Variation:** 21.9%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	ng/ml	4	M 12.859	9.7	0.78	1.32a			
2	No Result	ng/ml	8	M 155.279	2.2	1.49	12.93			
3	60.100	ng/ml	9	M 58.459	6.2	1.51	5.10a	0.32	119	2.81
4	241.500	ng/ml	10	M 249.009	3.6	3.54	20.74	-0.36	116	-3.02
5	34.700	ng/ml	9	M 35.174	2.9	0.43	2.93	-0.16	120	-1.35
6	132.000	ng/ml	8	M 130.225	2.3	1.32	10.85	0.16	120	1.36
7	232.400	ng/ml	7	M 219.239	5.4	5.57	19.09a	0.69	86	6.00
8	12.100	ng/ml	7	M 11.500	6.5	0.35	1.02a	0.59	92	5.22
9	110.100	ng/ml	8	M 110.463	2.4	1.16	9.20	-0.04	120	-0.33
10	56.600	ng/ml	10	M 57.403	8.6	1.95	5.16a	-0.16	120	-1.40
11	36.400	ng/ml	8	M 34.305	7.7	1.17	3.09a	0.68	85	6.11
12	220.900	ng/ml	10	M 211.090	4.7	3.93	17.58	0.56	97	4.65

	Cycle 16	Cycle 17
Cycle Average SDI	-0.02	0.23
Cycle Average TS	111	108
Cycle Average %DEV	-0.21	2.01
Cycle Average Absolute SDI	0.33	0.37
Cycle Average Absolute %DEV	3.01	3.22







Beta-2-Microglobulin, mg/l

Method: bioMerieux, VIDAS
Instrument: Biomerieux Vidas/miniVidas/Vidas 3
Reagent: bioMerieux

RIQAS TDPA: 15.6% **Biological Variation:** 9%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	mg/l	22	M 1.033	5.9	0.02	0.10			
2	No Result	mg/l	19	M 4.770	12.0	0.16	0.48a			
3	3.430	mg/l	32	M 3.265	10.5	0.08	0.31	0.53	99	5.04
4	13.430	mg/l	27	M 7.057	16.5	0.28	0.73a	8.78	10	90.32
5	2.580	mg/l	36	M 2.233	10.5	0.05	0.21	1.64	50	15.52
6	3.370	mg/l	30	M 4.446	12.6	0.13	0.44a	-2.44	31	-24.20
7	7.150	mg/l	20	M 7.374	11.8	0.24	0.74a	-0.30	120	-3.03
8	1.150	mg/l	28	M 1.083	8.7	0.02	0.10	0.65	90	6.17
9	4.340	mg/l	28	M 4.231	12.7	0.13	0.42a	0.26	120	2.57
10	3.270	mg/l	30	M 3.538	8.7	0.07	0.34	-0.80	81	-7.57
11	2.270	mg/l	30	M 2.405	7.5	0.04	0.23	-0.59	94	-5.60
12	10.230	mg/l	26	M 7.562	18.5	0.34	0.79a	3.36	15	35.28

	Cycle 16	Cycle 17
Cycle Average SDI	-0.21	1.11
Cycle Average TS	85	71
Cycle Average %DEV	-2.33	11.45
Cycle Average Absolute SDI	0.85	1.94
Cycle Average Absolute %DEV	9.15	19.53







CA 125, U/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse G Series
Reagent: Fujirebio Inc.

RIQAS TDPA: 12.1% **Biological Variation:** 35.4%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	U/ml	5	M 19.100	6.1	0.65	1.55a			
2	No Result	U/ml	8	M 138.763	6.8	4.18	11.03a			
3	96.400	U/ml	9	M 93.722	3.9	1.52	6.89	0.39	113	2.86
4	186.200	U/ml	10	M 190.170	8.9	6.70	15.51a	-0.26	120	-2.09
5	56.100	U/ml	11	M 54.955	7.9	1.63	4.36a	0.26	120	2.08
6	112.900	U/ml	8	M 113.063	1.5	0.75	8.32	-0.02	120	-0.14
7	208.500	U/ml	8	M 202.225	3.9	3.45	14.88	0.42	109	3.10
8	19.100	U/ml	8	M 18.189	10.5	0.84	1.58a	0.58	88	5.01
9	109.100	U/ml	9	M 108.157	14.1	6.34	10.17a	0.09	120	0.87
10	89.900	U/ml	9	M 96.944	5.5	2.22	7.47a	-0.94	72	-7.27
11	55.200	U/ml	9	M 55.982	7.6	1.78	4.49a	-0.17	120	-1.40
12	196.800	U/ml	8	M 198.846	4.5	3.94	14.63	-0.14	120	-1.03

	Cycle 16	Cycle 17
Cycle Average SDI	0.20	0.02
Cycle Average TS	110	110
Cycle Average %DEV	1.64	0.20
Cycle Average Absolute SDI	0.35	0.33
Cycle Average Absolute %DEV	2.79	2.59



CA 15-3, U/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse GSeries
Reagent: Fujirebio Inc.





RIQAS TDPA: 14% **Biological Variation:** 20.8%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	U/ml	3	M 26.267	2.4	0.46	2.24			
2	No Result	U/ml	7	M 67.171	5.7	1.82	6.00a			
3	51.900	U/ml	8	M 50.113	2.1	0.47	4.27	0.42	109	3.57
4	84.200	U/ml	8	M 85.250	4.4	1.67	7.26	-0.14	120	-1.23
5	36.800	U/ml	9	M 37.156	6.0	0.93	3.16	-0.11	120	-0.96
6	56.800	U/ml	9	M 57.256	8.1	1.94	5.25a	-0.09	120	-0.80
7	94.400	U/ml	6	M 93.083	6.9	3.29	8.58a	0.15	120	1.41
8	25.500	U/ml	7	M 25.067	8.3	0.98	2.35a	0.18	120	1.73
9	61.100	U/ml	8	M 60.178	9.1	2.43	5.67a	0.16	120	1.53
10	43.600	U/ml	7	M 48.610	3.0	0.69	4.14	-1.21	63	-10.31
11	37.300	U/ml	9	M 35.867	8.9	1.33	3.33a	0.43	104	4.00
12	102.400	U/ml	8	M 98.630	3.1	1.34	8.39	0.45	106	3.82

Cycle 16 Cycle 17

Cycle Average SDI 0.31 0.02
Cycle Average TS 105 110
Cycle Average %DEV 2.91 0.28

Cycle Average Absolute SDI 0.40 0.34
Cycle Average Absolute %DEV 3.66 2.94







CA I9-9, U/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse GSeries
Reagent: Fujirebio Inc.

RIQAS TDPA: 14.9% **Biological Variation:** 46.03%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	U/ml	4	M 36.250	21.8	4.93	5.92a			
2	No Result	U/ml	8	M 235.425	6.2	6.43	22.27a			
3	151.400	U/ml	8	M 147.950	4.9	3.19	13.40	0.26	120	2.33
4	296.500	U/ml	10	M 332.750	6.0	7.89	30.14	-1.20	64	-10.89
5	96.700	U/ml	9	M 93.367	4.2	1.63	8.46	0.39	112	3.57
6	205.200	U/ml	8	M 197.888	3.6	3.12	17.93	0.41	111	3.70
7	223.100	U/ml	6	M 220.352	11.1	12.53	23.57a	0.12	120	1.25
8	29.400	U/ml	5	M 31.400	1.4	0.24	2.84	-0.70	87	-6.37
9	129.200	U/ml	6	M 122.717	3.0	1.90	11.12	0.58	95	5.28
10	99.400	U/ml	8	M 101.865	6.4	2.86	9.66a	-0.26	120	-2.42
11	67.800	U/ml	7	M 64.229	3.8	1.16	5.82	0.61	93	5.56
12	244.000	U/ml	8	M 217.395	2.9	2.75	19.69	1.35	59	12.24

	Cycle 16	Cycle 17
Cycle Average SDI	0.15	0.16
Cycle Average TS	114	98
Cycle Average %DEV	1.47	1.42
Cycle Average Absolute SDI	0.25	0.59
Cycle Average Absolute %DEV	2.47	5.36







CEA, ng/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse G Series
Reagent: Fujirebio Inc.

RIQAS TDPA: 11.3% **Biological Variation:** 24.7%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	ng/ml	5	M 5.080	3.8	0.11	0.37a			
2	No Result	ng/ml	10	M 54.980	5.6	1.22	3.97a			
3	31.800	ng/ml	11	M 31.364	4.3	0.51	2.15	0.20	120	1.39
4	74.100	ng/ml	12	M 79.550	5.2	1.49	5.47	-1.00	72	-6.85
5	18.000	ng/ml	10	M 17.820	2.2	0.15	1.22	0.15	120	1.01
6	44.700	ng/ml	9	M 44.922	1.3	0.24	3.09	-0.07	120	-0.49
7	67.500	ng/ml	6	M 66.150	2.0	0.66	4.54	0.30	120	2.04
8	5.000	ng/ml	8	M 5.288	3.6	0.08	0.36	-0.79	82	-5.44
9	34.400	ng/ml	10	M 34.735	6.6	0.91	2.55a	-0.13	120	-0.96
10	30.300	ng/ml	11	M 30.075	6.9	0.78	2.21a	0.10	120	0.75
11	17.600	ng/ml	9	M 17.356	5.1	0.37	1.25a	0.20	120	1.41
12	68.800	ng/ml	10	M 66.575	4.3	1.14	4.57	0.49	103	3.34

	Cycle 16	Cycle 17
Cycle Average SDI	-0.16	-0.06
Cycle Average TS	116	110
Cycle Average %DEV	-1.35	-0.38
Cycle Average Absolute SDI	0.21	0.34
Cycle Average Absolute %DEV	1.69	2.37







Cortisol, ug/dl

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 14.3% **Biological Variation:** 22.8%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ug/dl	85	M 7.426	9.1	0.09	0.65			
2	N/A	ug/dl	112	M 32.539	6.4	0.25	2.83			
3	28.740	ug/dl	11	M 27.085	5.8	0.60	2.35	0.70	87	6.11
4	41.600	ug/dl	14	M 42.133	7.9	1.12	3.83a	-0.14	120	-1.27
5	20.100	ug/dl	13	M 18.523	6.1	0.39	1.61	0.98	72	8.52
6	21.470	ug/dl	18	M 25.914	9.4	0.72	2.36a	-1.88	42	-17.15
7	35.560	ug/dl	17	M 36.321	7.7	0.84	3.16	-0.24	120	-2.09
8	7.960	ug/dl	19	M 7.685	9.4	0.21	0.70a	0.39	110	3.58
9	22.820	ug/dl	21	M 21.937	6.2	0.37	1.91	0.46	105	4.03
10	24.020	ug/dl	25	M 22.812	7.5	0.43	1.98	0.61	93	5.29
11	14.450	ug/dl	26	M 16.015	7.6	0.30	1.39	-1.12	66	-9.77
12	38.370	ug/dl	22	M 36.605	5.8	0.56	3.18	0.55	97	4.82

	Cycle 16	Cycle 17
Cycle Average SDI	-0.43	0.03
Cycle Average TS	92	91
Cycle Average %DEV	-3.81	0.21
Cycle Average Absolute SDI	0.70	0.71
Cycle Average Absolute %DEV	6.26	6.26







DHEA-S, ug/dl

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 16.6% **Biological Variation:** 10.9%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ug/dl	30	M 75.021	13.9	2.39	7.93a			
2	N/A	ug/dl	33	M 491.059	7.4	7.92	49.56			
3	431.300	ug/dl	6	M 421.806	4.5	9.59	42.57	0.22	120	2.25
4	556.560	ug/dl	6	M 615.434	6.5	20.42	65.38a	-0.90	74	-9.57
5	212.760	ug/dl	7	M 223.285	5.7	5.97	22.53	-0.47	105	-4.71
6	329.000	ug/dl	8	M 332.769	6.5	9.54	33.58	-0.11	120	-1.13
7	676.000	ug/dl	8	M 663.487	5.7	16.70	66.96	0.19	120	1.89
8	61.570	ug/dl	7	M 69.024	8.8	2.88	7.53a	-0.99	69	-10.80
9	388.770	ug/dl	10	M 377.722	13.4	19.97	43.03a	0.26	120	2.92
10	388.740	ug/dl	9	M 384.392	4.9	7.79	38.79	0.11	120	1.13
11	228.450	ug/dl	9	M 223.057	3.7	3.45	22.51	0.24	120	2.42
12	652.950	ug/dl	9	M 659.926	8.9	24.34	70.90a	-0.10	120	-1.06

	Cycle 16	Cycle 17
Cycle Average SDI	0.24	-0.16
Cycle Average TS	94	109
Cycle Average %DEV	2.73	-1.67
Cycle Average Absolute SDI	0.71	0.36
Cycle Average Absolute %DEV	7.91	3.79







Digoxin, ng/ml

Method: bioMerieux, VIDAS
Instrument: Biomerieux Vidas/miniVidas/Vidas 3
Reagent: bioMerieux

RIQAS TDPA: 16.5% **Biological Variation:** N/A

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	ng/ml	5	M 0.526	5.1	0.02	0.05			
2	No Result	ng/ml	7	M 2.989	8.9	0.13	0.32a			
3	2.040	ng/ml	11	M 2.210	8.7	0.07	0.23a	-0.73	83	-7.69
4	3.510	ng/ml	9	M 3.797	10.9	0.17	0.42a	-0.69	84	-7.55
5	1.120	ng/ml	9	M 1.228	10.2	0.05	0.13a	-0.81	77	-8.78
6	2.410	ng/ml	7	M 2.140	1.6	0.02	0.21	1.26	62	12.62
7	3.900	ng/ml	7	M 3.647	4.9	0.08	0.37	0.69	88	6.93
8	0.370	ng/ml	9	M 0.436	13.3	0.02	0.05a	-1.32	54	-15.05
9	1.750	ng/ml	8	M 1.921	6.7	0.06	0.19	-0.89	77	-8.91
10	2.030	ng/ml	7	M 2.113	6.5	0.07	0.22a	-0.37	112	-3.92
11	1.340	ng/ml	7	M 1.223	6.5	0.04	0.13a	0.91	74	9.58
12	4.010	ng/ml	8	M 3.704	5.4	0.09	0.37	0.82	80	8.27

	Cycle 16	Cycle 17
Cycle Average SDI	0.54	-0.11
Cycle Average TS	85	79
Cycle Average %DEV	6.29	-1.45
Cycle Average Absolute SDI	0.85	0.85
Cycle Average Absolute %DEV	9.76	8.93







Ferritin, ng/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 15.1% **Biological Variation:** 16.9%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/ml	145	M 13.600	7.7	0.11	1.25			
2	N/A	ng/ml	189	M 204.449	5.4	1.01	18.77			
3	92.100	ng/ml	17	M 88.016	5.2	1.39	8.08	0.51	101	4.64
4	349.300	ng/ml	21	M 320.359	6.8	5.91	29.41	0.98	72	9.03
5	45.800	ng/ml	22	M 49.176	3.7	0.49	4.51	-0.75	84	-6.87
6	176.900	ng/ml	22	M 159.015	2.9	1.25	14.60	1.23	63	11.25
7	327.500	ng/ml	22	M 311.576	3.7	3.08	28.60	0.56	97	5.11
8	14.200	ng/ml	25	M 14.166	8.1	0.29	1.30	0.03	120	0.24
9	162.800	ng/ml	26	M 157.882	4.3	1.68	14.49	0.34	119	3.11
10	100.300	ng/ml	30	M 95.802	4.6	1.00	8.79	0.51	101	4.69
11	53.700	ng/ml	29	M 53.416	4.1	0.50	4.90	0.06	120	0.53
12	353.600	ng/ml	31	M 318.786	5.4	3.87	29.27	1.19	64	10.92

	Cycle 16	Cycle 17
Cycle Average SDI	0.20	0.46
Cycle Average TS	101	94
Cycle Average %DEV	1.73	4.27
Cycle Average Absolute SDI	0.52	0.61
Cycle Average Absolute %DEV	4.48	5.64







Folate, ng/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 18.9% **Biological Variation:** 39%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/ml	10	M 3.212	25.4	0.32	0.49a			
2	N/A	ng/ml	10	M 11.451	12.2	0.55	1.43a			
3	10.260	ng/ml	11	M 10.655	6.0	0.24	1.22	-0.32	120	-3.71
4	17.290	ng/ml	14	M 16.473	10.9	0.60	1.98a	0.41	108	4.96
5	6.900	ng/ml	13	M 7.318	8.0	0.20	0.84	-0.50	102	-5.71
6	11.080	ng/ml	16	M 11.026	10.7	0.37	1.27	0.04	120	0.49
7	No Result	ng/ml	15	M 13.913	7.4	0.33	1.60			
8	No Result	ng/ml	17	M 5.243	11.2	0.18	0.60			
9	10.180	ng/ml	22	M 10.182	13.2	0.36	1.22a	-0.00	120	-0.02
10	No Result	ng/ml	22	M 10.251	10.6	0.29	1.18			
11	7.100	ng/ml	21	M 7.622	9.9	0.21	0.88	-0.60	94	-6.84
12	12.830	ng/ml	20	M 13.885	10.9	0.42	1.60	-0.66	90	-7.60

	Cycle 16	Cycle 17
Cycle Average SDI	0.68	-0.23
Cycle Average TS	67	108
Cycle Average %DEV	8.75	-2.63
Cycle Average Absolute SDI	1.22	0.36
Cycle Average Absolute %DEV	16.10	4.19







FSH, mU/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse G Series
Reagent: Fujirebio Inc.

RIQAS TDPA: 11.7% **Biological Variation:** 21.19%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	mU/ml	3	M 5.067	11.6	0.42	0.56a			
2	No Result	mU/ml	9	M 35.978	3.4	0.51	2.56			
3	23.100	mU/ml	10	M 23.060	4.9	0.45	1.64	0.02	120	0.17
4	49.100	mU/ml	11	M 49.845	3.7	0.70	3.55	-0.21	120	-1.50
5	14.300	mU/ml	9	M 13.778	3.1	0.18	0.98	0.53	99	3.79
6	28.600	mU/ml	10	M 28.150	6.8	0.76	2.14a	0.21	120	1.60
7	51.500	mU/ml	7	M 51.214	1.4	0.35	3.64	0.08	120	0.56
8	5.400	mU/ml	9	M 5.422	4.7	0.11	0.39	-0.06	120	-0.41
9	27.800	mU/ml	11	M 27.518	4.7	0.49	1.96	0.14	120	1.02
10	23.800	mU/ml	10	M 24.515	4.4	0.42	1.74	-0.41	110	-2.92
11	15.200	mU/ml	11	M 14.724	2.5	0.14	1.05	0.45	106	3.24
12	49.500	mU/ml	12	M 50.181	3.4	0.61	3.57	-0.19	120	-1.36

	Cycle 16	Cycle 17
Cycle Average SDI	-0.41	0.06
Cycle Average TS	104	116
Cycle Average %DEV	-3.20	0.42
Cycle Average Absolute SDI	0.44	0.23
Cycle Average Absolute %DEV	3.43	1.66







hCG, mU/ml

Method: Siemens Atellica IM
Instrument: Siemens ADVIA Centaur XP/XPT/Classic
Reagent: Siemens

RIQAS TDPA: 13.2% **Biological Variation:** N/A

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	mU/ml	142	M 19.105	6.7	0.13	1.53			
2	N/A	mU/ml	189	M 168.660	4.0	0.61	13.54			
3	26.700	mU/ml	13	M 26.219	5.3	0.48	2.10	0.23	120	1.83
4	339.800	mU/ml	16	M 316.962	4.3	4.26	25.44	0.90	76	7.21
5	21.800	mU/ml	16	M 21.256	7.7	0.51	1.71	0.32	120	2.56
6	193.400	mU/ml	17	M 186.501	5.1	2.88	14.97	0.46	105	3.70
7	118.800	mU/ml	17	M 115.878	3.4	1.19	9.30	0.31	120	2.52
8	11.100	mU/ml	23	M 10.445	5.7	0.16	0.84	0.78	82	6.27
9	76.200	mU/ml	24	M 75.522	4.6	0.89	6.06	0.11	120	0.90
10	12.000	mU/ml	31	M 11.467	5.3	0.14	0.92	0.58	95	4.65
11	No Result	mU/ml	28	M 11.306	6.6	0.18	0.91			
12	116.400	mU/ml	27	M 113.332	3.4	0.92	9.09	0.34	119	2.71

	Cycle 16	Cycle 17
Cycle Average SDI	0.05	0.45
Cycle Average TS	92	106
Cycle Average %DEV	0.40	3.59
Cycle Average Absolute SDI	0.67	0.45
Cycle Average Absolute %DEV	5.63	3.59







IgE, U/ml

Method: Siemens Centaur XP/XPT/Classic
Instrument: Siemens ADVIA Centaur XP/XPT/Classic
Reagent: Siemens

RIQAS TDPA: 14.7% **Biological Variation:** N/A

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	U/ml	59	M 234.696	6.0	2.29	20.97			
2	No Result	U/ml	80	M 310.655	7.1	3.08	27.76			
3	108.900	U/ml	83	M 101.615	5.9	0.83	9.08	0.80	81	7.17
4	583.100	U/ml	86	M 520.286	7.2	5.08	46.50	1.35	59	12.07
5	197.100	U/ml	82	M 168.751	6.3	1.48	15.08	1.88	44	16.80
6	406.400	U/ml	85	M 384.623	6.2	3.22	34.37	0.63	91	5.66
7	602.700	U/ml	59	M 620.181	6.3	6.39	55.43	-0.32	120	-2.82
8	302.400	U/ml	65	M 274.722	5.3	2.27	24.55	1.13	66	10.07
9	421.000	U/ml	75	M 436.996	5.4	3.42	39.05	-0.41	110	-3.66
10	102.400	U/ml	69	M 108.387	6.1	0.99	9.69	-0.62	92	-5.52
11	172.300	U/ml	81	M 189.397	5.3	1.39	16.93	-1.01	71	-9.03
12	627.300	U/ml	64	M 586.690	7.9	7.21	52.43	0.77	83	6.92

	Cycle 16	Cycle 17
Cycle Average SDI	-0.27	0.42
Cycle Average TS	87	82
Cycle Average %DEV	-2.29	3.77
Cycle Average Absolute SDI	0.80	0.89
Cycle Average Absolute %DEV	6.80	7.97







LH, mU/ml

Method: Siemens Centaur XP/XPT/Classic
Instrument: Siemens ADVIA Centaur XP/XPT/Classic
Reagent: Siemens

RIQAS TDPA: 13% **Biological Variation:** 27.92%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	mU/ml	126	M 2.579	5.9	0.02	0.20			
2	No Result	mU/ml	169	M 42.473	5.2	0.21	3.36			
3	37.410	mU/ml	178	M 34.254	5.5	0.18	2.71	1.17	65	9.21
4	55.420	mU/ml	178	M 51.105	5.5	0.26	4.04	1.07	69	8.44
5	18.060	mU/ml	167	M 17.529	6.1	0.10	1.39	0.38	113	3.03
6	27.390	mU/ml	166	M 28.814	4.9	0.14	2.28	-0.63	92	-4.94
7	51.880	mU/ml	129	M 52.744	5.5	0.32	4.17	-0.21	120	-1.64
8	3.040	mU/ml	138	M 3.085	6.2	0.02	0.24	-0.18	120	-1.46
9	29.680	mU/ml	170	M 29.183	4.1	0.11	2.31	0.22	120	1.70
10	36.020	mU/ml	152	M 35.201	5.1	0.18	2.78	0.29	120	2.33
11	19.800	mU/ml	169	M 19.329	4.8	0.09	1.53	0.31	120	2.44
12	52.960	mU/ml	141	M 50.928	5.4	0.29	4.03	0.50	101	3.99

	Cycle 16	Cycle 17
Cycle Average SDI	0.22	0.29
Cycle Average TS	102	104
Cycle Average %DEV	1.58	2.31
Cycle Average Absolute SDI	0.52	0.50
Cycle Average Absolute %DEV	3.76	3.92



Oestradiol, pg/ml

Method: Fujirebio Lumipulse G Series

Instrument: Fujirebio Lumipulse GSeries





Reagent: Fujirebio Inc.

RIQAS TDPA: 17.9%

Biological Variation: 26.86%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	pg/ml	3	M 49.822	25.5	9.16	10.65a			
2	No Result	pg/ml	5	M 482.844	3.4	9.30	52.55			
3	301.100	pg/ml	6	M 315.123	3.9	6.27	34.29	-0.41	110	-4.45
4	615.400	pg/ml	5	M 648.140	5.6	20.24	70.53	-0.46	105	-5.05
5	154.200	pg/ml	5	M 162.620	6.2	5.62	18.57a	-0.45	104	-5.18
6	310.600	pg/ml	5	M 319.480	4.1	7.27	34.77	-0.26	120	-2.78
7	496.600	pg/ml	6	M 554.012	10.9	30.74	67.68a	-0.85	74	-10.36
8	53.900	pg/ml	5	M 45.040	14.2	3.58	6.07a	1.46	46	19.67
9	226.900	pg/ml	6	M 273.378	20.2	28.17	40.97a	-1.13	52	-17.00
10	233.400	pg/ml	6	M 255.174	17.6	22.86	35.97a	-0.61	82	-8.53
11	130.200	pg/ml	7	M 140.839	9.5	6.34	16.59a	-0.64	87	-7.55
12	495.200	pg/ml	6	M 528.648	4.7	12.74	57.53	-0.58	95	-6.33

	Cycle 16	Cycle 17
Cycle Average SDI	-0.39	-0.39
Cycle Average TS	97	88
Cycle Average %DEV	-4.27	-4.76
Cycle Average Absolute SDI	0.56	0.69
Cycle Average Absolute %DEV	6.11	8.69







Progesterone, ng/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 18.8% **Biological Variation:** N/A

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/ml	101	M 0.717	15.3	0.01	0.08			
2	N/A	ng/ml	126	M 15.273	6.6	0.11	1.75			
3	8.210	ng/ml	9	M 8.449	2.5	0.09	0.97	-0.25	120	-2.82
4	24.540	ng/ml	13	M 25.267	6.6	0.58	2.89	-0.25	120	-2.88
5	4.180	ng/ml	12	M 4.181	5.8	0.09	0.48	-0.00	120	-0.03
6	10.770	ng/ml	13	M 10.885	3.1	0.12	1.24	-0.09	120	-1.06
7	25.390	ng/ml	15	M 24.830	5.6	0.45	2.84	0.20	120	2.25
8	0.930	ng/ml	15	M 1.012	9.4	0.03	0.12	-0.71	86	-8.11
9	11.390	ng/ml	19	M 12.072	6.7	0.23	1.38	-0.49	102	-5.65
10	9.180	ng/ml	20	M 8.710	4.0	0.10	1.00	0.47	104	5.39
11	4.030	ng/ml	21	M 4.639	3.6	0.04	0.53	-1.15	66	-13.12
12	23.450	ng/ml	19	M 23.860	3.2	0.22	2.73	-0.15	120	-1.72

	Cycle 16	Cycle 17
Cycle Average SDI	0.33	-0.24
Cycle Average TS	97	108
Cycle Average %DEV	3.62	-2.77
Cycle Average Absolute SDI	0.64	0.38
Cycle Average Absolute %DEV	7.05	4.30







Prolactin, uU/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 18.6% **Biological Variation:** 29.4%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	uU/ml	122	M 139.857	3.7	0.59	15.82			
2	N/A	uU/ml	156	M 738.018	5.1	3.78	83.46			
3	509.776	uU/ml	8	M 494.446	5.6	12.25	55.91	0.27	120	3.10
4	963.540	uU/ml	12	M 922.907	4.4	14.80	104.36	0.39	113	4.40
5	296.376	uU/ml	11	M 299.920	3.6	4.12	33.91	-0.10	120	-1.18
6	23.390	uU/ml	14	M 508.665	4.3	7.26	57.52	-8.44	10	-95.40
7	1107.064	uU/ml	17	M 1046.034	3.9	12.36	118.29	0.52	100	5.83
8	162.820	uU/ml	19	M 162.202	3.3	1.55	18.34	0.03	120	0.38
9	645.752	uU/ml	16	M 631.333	4.6	9.05	71.39	0.20	120	2.28
10	631.124	uU/ml	22	M 623.955	3.8	6.34	70.56	0.10	120	1.15
11	385.628	uU/ml	21	M 385.258	6.0	6.26	43.57	0.01	120	0.10
12	1089.892	uU/ml	23	M 1063.487	4.9	13.69	120.26	0.22	120	2.48

	Cycle 16	Cycle 17
Cycle Average SDI	0.09	-0.68
Cycle Average TS	115	106
Cycle Average %DEV	1.26	-7.69
Cycle Average Absolute SDI	0.30	1.03
Cycle Average Absolute %DEV	4.29	11.63







PSA, Free, ng/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 12.2% **Biological Variation:** N/A

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/ml	82	M 1.142	4.4	0.01	0.08			
2	N/A	ng/ml	119	M 13.528	4.9	0.08	1.00			
3	8.100	ng/ml	11	M 8.273	5.8	0.18	0.61	-0.28	120	-2.09
4	20.540	ng/ml	12	M 20.476	6.8	0.50	1.60a	0.04	120	0.31
5	4.950	ng/ml	15	M 4.520	6.3	0.09	0.34	1.28	61	9.51
6	11.380	ng/ml	15	M 10.445	5.5	0.19	0.77	1.21	63	8.96
7	20.670	ng/ml	15	M 19.411	4.6	0.29	1.44	0.87	77	6.48
8	1.200	ng/ml	16	M 1.162	3.6	0.01	0.09	0.44	107	3.28
9	9.290	ng/ml	16	M 10.295	6.0	0.19	0.76	-1.32	60	-9.76
10	9.450	ng/ml	20	M 8.751	4.6	0.11	0.65	1.08	68	7.99
11	5.100	ng/ml	17	M 4.966	4.0	0.06	0.37	0.36	115	2.70
12	21.830	ng/ml	17	M 19.415	2.4	0.14	1.44	1.68	49	12.44

	Cycle 16	Cycle 17
Cycle Average SDI	0.10	0.54
Cycle Average TS	89	84
Cycle Average %DEV	0.70	3.98
Cycle Average Absolute SDI	0.73	0.86
Cycle Average Absolute %DEV	5.22	6.35







PSA, Total, ng/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 12.3% **Biological Variation:** 33.6%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/ml	160	M 1.465	4.6	0.01	0.11			
2	N/A	ng/ml	213	M 18.384	5.1	0.08	1.37			
3	11.420	ng/ml	13	M 11.451	3.4	0.13	0.86	-0.04	120	-0.27
4	28.270	ng/ml	16	M 27.951	5.1	0.45	2.09	0.15	120	1.14
5	6.110	ng/ml	21	M 6.057	4.7	0.08	0.45	0.12	120	0.87
6	15.890	ng/ml	20	M 14.368	4.0	0.16	1.07	1.42	56	10.60
7	26.630	ng/ml	21	M 26.310	3.5	0.25	1.97	0.16	120	1.22
8	1.570	ng/ml	24	M 1.568	3.0	0.01	0.12	0.02	120	0.14
9	No Result	ng/ml	24	M 13.833	4.9	0.17	1.03			
10	11.910	ng/ml	29	M 12.309	4.8	0.14	0.92	-0.43	108	-3.24
11	7.300	ng/ml	30	M 6.834	4.0	0.06	0.51	0.91	76	6.83
12	No Result	ng/ml	28	M 26.244	5.0	0.31	1.96			

	Cycle 16	Cycle 17
Cycle Average SDI	0.29	0.29
Cycle Average TS	85	105
Cycle Average %DEV	2.16	2.16
Cycle Average Absolute SDI	0.88	0.41
Cycle Average Absolute %DEV	6.59	3.04







Parathyroid hormone, pg/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: N/A **Biological Variation:** 30.2%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	pg/ml	7	M 1.886	30.0	0.27	0.62a		N/A	
2	N/A	pg/ml	17	M 655.171	17.3	34.32	118.29a		N/A	
3	314.300	pg/ml	10	M 260.350	13.2	13.63	37.07a	1.46	N/A	20.72
4	1630.000	pg/ml	9	M 1575.697	9.1	59.42	154.48a	0.35	N/A	3.45
5	2.000	pg/ml	4	M 2.125	27.3	0.36	0.68a	-0.18	N/A	-5.88
6	1.600	pg/ml	5	M 4.140	11.4	0.26	0.54a	-4.71	N/A	-61.35
7	No Result	pg/ml	12	M 1255.650	5.4	24.53	72.27a		N/A	
8	61.700	pg/ml	12	M 46.921	29.8	5.05	14.89a	0.99	N/A	31.50
9	465.900	pg/ml	14	M 420.706	11.4	15.98	50.42a	0.90	N/A	10.74
10	488.000	pg/ml	15	M 454.644	5.9	8.60	28.01a	1.19	N/A	7.34
11	198.400	pg/ml	18	M 172.527	11.6	5.92	20.10	1.29	N/A	15.00
12	No Result	pg/ml	18	M 1266.932	6.1	22.62	76.77		N/A	

	Cycle 16	Cycle 17
Cycle Average SDI	-1.23	0.16
Cycle Average TS	N/A	N/A
Cycle Average %DEV	-14.20	2.69
Cycle Average Absolute SDI	1.58	1.38
Cycle Average Absolute %DEV	17.03	19.50







Free T3, pmol/l

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 15.4% **Biological Variation:** 11.3%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	pmol/l	174	M 3.604	3.9	0.01	0.34			
2	N/A	pmol/l	227	M 18.701	5.2	0.08	1.75			
3	13.490	pmol/l	15	M 13.384	2.5	0.11	1.25	0.08	120	0.79
4	26.660	pmol/l	19	M 27.286	5.8	0.45	2.55	-0.25	120	-2.30
5	9.490	pmol/l	20	M 9.155	2.1	0.05	0.86	0.39	112	3.66
6	15.790	pmol/l	24	M 15.248	2.6	0.10	1.43	0.38	114	3.56
7	26.100	pmol/l	22	M 24.399	4.8	0.31	2.28	0.74	84	6.97
8	3.950	pmol/l	29	M 3.894	3.5	0.03	0.36	0.15	120	1.43
9	14.650	pmol/l	30	M 14.544	3.4	0.11	1.36	0.08	120	0.73
10	12.670	pmol/l	31	M 12.718	2.4	0.07	1.19	-0.04	120	-0.38
11	9.530	pmol/l	37	M 9.295	2.6	0.05	0.87	0.27	120	2.53
12	23.090	pmol/l	35	M 25.486	7.1	0.38	2.39	-1.00	71	-9.40

	Cycle 16	Cycle 17
Cycle Average SDI	-0.05	0.08
Cycle Average TS	109	110
Cycle Average %DEV	-0.42	0.76
Cycle Average Absolute SDI	0.41	0.34
Cycle Average Absolute %DEV	3.51	3.17







Free T4, pmol/l

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 12.3% **Biological Variation:** 8%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	pmol/l	200	M 8.197	7.9	0.06	0.61			
2	N/A	pmol/l	255	M 40.272	6.1	0.19	3.01			
3	32.660	pmol/l	17	M 32.495	3.1	0.30	2.43	0.07	120	0.51
4	55.760	pmol/l	21	M 60.408	5.3	0.87	4.52	-1.03	70	-7.69
5	21.310	pmol/l	23	M 20.006	4.7	0.24	1.50	0.87	78	6.52
6	30.230	pmol/l	27	M 32.104	4.4	0.34	2.40	-0.78	82	-5.84
7	56.340	pmol/l	25	M 59.200	5.0	0.73	4.43	-0.65	91	-4.83
8	8.730	pmol/l	29	M 8.739	7.2	0.15	0.65	-0.01	120	-0.11
9	29.180	pmol/l	35	M 29.104	4.7	0.29	2.18	0.03	120	0.26
10	30.520	pmol/l	37	M 31.125	4.0	0.25	2.33	-0.26	120	-1.94
11	20.930	pmol/l	36	M 20.394	3.9	0.17	1.53	0.35	117	2.63
12	53.720	pmol/l	35	M 56.134	3.9	0.46	4.20	-0.57	96	-4.30

	Cycle 16	Cycle 17
Cycle Average SDI	-0.46	-0.20
Cycle Average TS	94	101
Cycle Average %DEV	-3.53	-1.48
Cycle Average Absolute SDI	0.67	0.46
Cycle Average Absolute %DEV	5.11	3.46







Testosterone, ng/dl

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 17.9% **Biological Variation:** 13.61%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	ng/dl	41	M 46.232	7.0	0.63	5.03			
2	N/A	ng/dl	56	M 604.693	9.3	9.39	65.80			
3	424.840	ng/dl	7	M 437.498	2.1	4.40	47.61	-0.27	120	-2.89
4	730.770	ng/dl	10	M 723.378	2.2	6.42	78.72	0.09	120	1.02
5	231.240	ng/dl	11	M 245.069	4.6	4.21	26.67	-0.52	100	-5.64
6	439.910	ng/dl	13	M 417.536	4.2	6.03	45.44	0.49	102	5.36
7	690.030	ng/dl	13	M 688.715	5.5	13.07	74.95	0.02	120	0.19
8	38.020	ng/dl	16	M 40.019	3.6	0.45	4.35	-0.46	105	-4.99
9	326.870	ng/dl	21	M 381.559	9.2	9.61	41.52	-1.32	60	-14.33
10	360.830	ng/dl	21	M 367.738	5.7	5.67	40.02	-0.17	120	-1.88
11	No Result	ng/dl	21	M 202.727	4.1	2.27	22.06			
12	No Result	ng/dl	18	M 695.489	6.5	13.23	75.69			

	Cycle 16	Cycle 17
Cycle Average SDI	-0.38	-0.27
Cycle Average TS	96	106
Cycle Average %DEV	-4.64	-2.90
Cycle Average Absolute SDI	0.64	0.42
Cycle Average Absolute %DEV	7.77	4.54







Thyroglobulin, ng/ml

Method: Tosoh AIA Series
Instrument: TOSOH AIA Series
Reagent: TOSOH

RIQAS TDPA: 23.7% **Biological Variation:** 21.9%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	ng/ml	10	M 3.150	9.8	0.12	0.45			
2	No Result	ng/ml	11	M 51.445	7.1	1.39	7.41			
3	No Result	ng/ml	14	M 34.579	8.6	0.99	4.98			
4	No Result	ng/ml	14	M 67.207	9.2	2.06	9.68			
5	No Result	ng/ml	13	M 18.777	5.3	0.34	2.71			
6	No Result	ng/ml	13	M 37.277	6.7	0.86	5.37			
7	No Result	ng/ml	12	M 62.325	5.9	1.34	8.98			
8	No Result	ng/ml	8	M 1.825	8.2	0.07	0.26			
9	No Result	ng/ml	13	M 34.185	5.9	0.70	4.93			
10	No Result	ng/ml	11	M 33.691	4.7	0.59	4.85			
11	No Result	ng/ml	12	M 17.917	7.6	0.49	2.58			
12	No Result	ng/ml	8	M 63.975	2.4	0.68	9.22			

	Cycle 16	Cycle 17
Cycle Average SDI	0.01	N/A
Cycle Average TS	110	N/A
Cycle Average %DEV	0.15	N/A
Cycle Average Absolute SDI	0.29	N/A
Cycle Average Absolute %DEV	4.31	N/A







TSH, uU/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 12.5% **Biological Variation:** 23.7%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	uU/ml	101	M 0.127	4.1	0.00	0.01			
2	N/A	uU/ml	119	M 8.773	3.5	0.04	0.67			
3	2.678	uU/ml	19	M 2.539	4.8	0.03	0.19	0.72	86	5.46
4	23.216	uU/ml	22	M 22.901	4.0	0.24	1.74	0.18	120	1.37
5	1.316	uU/ml	24	M 1.273	5.0	0.02	0.10	0.44	107	3.34
6	7.530	uU/ml	28	M 7.513	5.5	0.10	0.57	0.03	120	0.23
7	21.340	uU/ml	26	M 20.484	5.3	0.27	1.56	0.55	98	4.18
8	0.104	uU/ml	31	M 0.100	2.9	0.00	0.01	0.47	105	3.53
9	7.061	uU/ml	33	M 7.103	3.1	0.05	0.54	-0.08	120	-0.58
10	2.230	uU/ml	42	M 2.137	5.3	0.02	0.16	0.58	96	4.37
11	1.155	uU/ml	39	M 1.160	3.5	0.01	0.09	-0.06	120	-0.42
12	21.535	uU/ml	39	M 21.108	4.3	0.18	1.60	0.27	120	2.02

	Cycle 16	Cycle 17
Cycle Average SDI	0.21	0.31
Cycle Average TS	113	109
Cycle Average %DEV	1.58	2.35
Cycle Average Absolute SDI	0.35	0.34
Cycle Average Absolute %DEV	2.62	2.55







Vitamin B12, pg/ml

Method: Siemens Atellica IM
Instrument: Siemens Atellica Solution
Reagent: Siemens

RIQAS TDPA: 18% **Biological Variation:** 30%

Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	N/A	pg/ml	10	M 199.314	15.0	11.79	24.79a			
2	N/A	pg/ml	11	M 1424.555	5.7	30.65	155.89			
3	668.000	pg/ml	11	M 703.724	6.2	16.46	77.01	-0.46	105	-5.08
4	1274.000	pg/ml	13	M 1225.839	5.0	21.04	134.15	0.36	116	3.93
5	419.000	pg/ml	15	M 405.210	10.9	14.31	46.59a	0.30	120	3.40
6	711.000	pg/ml	17	M 700.755	7.8	16.54	76.69	0.13	120	1.46
7	1225.000	pg/ml	16	M 1206.290	6.1	22.86	132.01	0.14	120	1.55
8	171.000	pg/ml	20	M 188.576	10.7	5.64	20.64	-0.85	79	-9.32
9	669.000	pg/ml	23	M 702.441	12.0	21.95	76.87	-0.44	108	-4.76
10	No Result	pg/ml	24	M 678.974	5.9	10.21	74.30			
11	322.000	pg/ml	25	M 426.214	9.8	10.42	46.64	-2.23	37	-24.45
12	987.000	pg/ml	24	M 1178.998	8.3	24.90	129.02	-1.49	54	-16.28

	Cycle 16	Cycle 17
Cycle Average SDI	0.53	-0.50
Cycle Average TS	100	95
Cycle Average %DEV	5.83	-5.51
Cycle Average Absolute SDI	0.53	0.71
Cycle Average Absolute %DEV	5.83	7.80







25-OH-Vitamin D, ng/ml

Method: Fujirebio Lumipulse G Series
Instrument: Fujirebio Lumipulse G Series
Reagent: Fujirebio Inc.

RIQAS TDPA: 27.1% **Biological Variation:** N/A


Sample	Result	Unit	N	Mean for Comparison	CV%	Um	SDPA	SDI	TS	%Deviation
1	No Result	ng/ml	3	M 8.133	23.6	1.39	1.93a			
2	No Result	ng/ml	7	M 30.914	1.6	0.24	5.09			
3	22.200	ng/ml	9	M 20.467	5.3	0.45	3.37	0.51	100	8.47
4	38.400	ng/ml	9	M 39.978	4.5	0.75	6.59	-0.24	120	-3.95
5	16.100	ng/ml	10	M 14.800	6.4	0.38	2.44	0.53	99	8.78
6	No Result	ng/ml	9	M 24.933	7.7	0.80	4.11			
7	37.000	ng/ml	9	M 36.744	5.9	0.91	6.05	0.04	120	0.70
8	4.800	ng/ml	10	M 6.540	27.5	0.71	1.29a	-1.35	51	-26.61
9	23.000	ng/ml	10	M 22.520	5.8	0.52	3.71	0.13	120	2.13
10	25.100	ng/ml	9	M 24.978	3.0	0.32	4.12	0.03	120	0.49
11	15.800	ng/ml	8	M 16.416	5.5	0.40	2.70	-0.23	120	-3.75
12	34.800	ng/ml	11	M 36.219	6.3	0.86	5.97	-0.24	120	-3.92

	Cycle 16	Cycle 17
Cycle Average SDI	0.43	-0.09
Cycle Average TS	96	108
Cycle Average %DEV	7.93	-1.96
Cycle Average Absolute SDI	0.71	0.37
Cycle Average Absolute %DEV	13.02	6.53



Cycle Average Absolute SDI

L Laboratory
C Country
W World



Cycle Average Absolute SDI

L	Laboratory
C	Country
W	World


Parameter

Cycle 16 ○

● Better
● Worse

Cycle 17

Acceptable performance: SDI <= 2



END OF REPORT

CERTIFICATE OF ACCEPTABLE PERFORMANCE

Biolab SRL
Via Igea 12C
ROMA
00135
ITALY

LABORATORY REF. NO. 260037/E

MONTHLY IMMUNOASSAY - CYCLE 17

30/12/2019

This is to certify that the above participant took part in a cycle of external quality assessment and achieved an acceptable level of performance (Cycle Average Absolute SDI ≤ 2) for the following parameters:

	Cycle Average Absolute SDI
AFP - Fujirebio Lumipulse G Series	0.37
Beta-2-Microglobulin - bioMerieux, VIDAS	1.94
CA 125 - Fujirebio Lumipulse G Series	0.33
CA 15-3 - Fujirebio Lumipulse G Series	0.34
CA 19-9 - Fujirebio Lumipulse G Series	0.59
CEA - Fujirebio Lumipulse G Series	0.34
Cortisol - Siemens Atellica IM	0.71
DHEA-S - Siemens Atellica IM	0.36
Digoxin - bioMerieux, VIDAS	0.85
Ferritin - Siemens Atellica IM	0.61
Folate - Siemens Atellica IM	0.36
FSH - Fujirebio Lumipulse G Series	0.23
hCG - Siemens Atellica IM	0.45
IgE - Siemens Centaur XP/XPT/Classic	0.89
LH - Siemens Centaur XP/XPT/Classic	0.50
Oestradiol - Fujirebio Lumipulse G Series	0.69
Progesterone - Siemens Atellica IM	0.38
Prolactin - Siemens Atellica IM	1.03
PSA, Free - Siemens Atellica IM	0.86
PSA, Total - Siemens Atellica IM	0.41
Parathyroid hormone - Siemens Atellica IM	1.38
Free T3 - Siemens Atellica IM	0.34
Free T4 - Siemens Atellica IM	0.46
Testosterone - Siemens Atellica IM	0.42
TSH - Siemens Atellica IM	0.34
Vitamin B12 - Siemens Atellica IM	0.71
25-OH-Vitamin D - Fujirebio Lumipulse G Series	0.37

This certificate is valid until 24/01/2021

Authorised by:
Stephen Doherty, RIQAS Manager

Page 1 of 1